



Evodial  
DIALYSER

## ARTICLES

---

Results of the HepZero study comparing heparin-grafted membrane and standard care show that heparin-grafted dialyser is safe and easy to use for heparin-free dialysis<sup>1</sup>

[https://www.kidney-international.org/article/S0085-2538\(15\)30458-0/pdf](https://www.kidney-international.org/article/S0085-2538(15)30458-0/pdf)

Anticoagulation in Chronic Haemodialysis: Progress Toward an Optimal Approach<sup>2</sup>

<https://onlinelibrary.wiley.com/doi/full/10.1111/sdi.12380>

AN69: Evolution of the World's First High Permeability Membrane<sup>3</sup>

<https://www.karger.com/Article/Abstract/328961>

The clinical evaluation of low-dose heparin in haemodialysis: a prospective study using the heparin-coated AN69 ST membrane<sup>4</sup>

<https://academic.oup.com/ndt/article/23/6/2003/1875941>

A noninferiority trial comparing a heparin-grafted membrane plus citrate-containing dialysate versus regional citrate anticoagulation: results of the CiTED study<sup>5</sup>

<https://academic.oup.com/ndt/article/32/4/707/3604955>

**Australia**  
Baxter Healthcare Ltd  
1 Baxter Drive, Old Toongabbie  
NSW 2146, Australia  
Tel: 1800 BAXTER (1800 229 837)

**New Zealand**  
Baxter Healthcare Ltd  
33 Vestey Drive, Mount Wellington  
Auckland 1060, New Zealand  
Tel: 0800 BAXTER (0800 229 837)

Baxter and Evodial are trademarks of Baxter International Inc. or its subsidiaries.

# Evidence Summary

## REFERENCES

---

<sup>1</sup>Laville, Maurice & Dorval, Marc & Fort, et al. (2014). Results of the HepZero study comparing heparin-grafted membrane and standard care show that heparin-grafted dialyser is safe and easy to use for heparin-free dialysis. *Kidney international*. 86. [10.1038/ki.2014.225](https://doi.org/10.1038/ki.2014.225)

<sup>2</sup>Kessler, M. , Moureau, F. and Nguyen, P. (2015), Anticoagulation in Chronic Hemodialysis: Progress Toward an Optimal Approach. *Semin Dial*, 28: 474-489. doi:[10.1111/sdi.12380](https://doi.org/10.1111/sdi.12380)

<sup>3</sup>Saito A, Kawanishi H, Yamashita AC, Mineshima M (eds): High-Performance Membrane Dialysers. *Contrib Nephrol*. Basel, Karger, 2011, vol 173, pp 119–129. doi: [10.1159/000328961](https://doi.org/10.1159/000328961)

<sup>4</sup>Jacques Chanard, Sylvie Lavaud, Hervé Maheut, Isabelle Kazes, Fabien Vitry, Philippe Rieu, The clinical evaluation of low-dose heparin in haemodialysis: a prospective study using the heparin-coated AN69 ST membrane, *Nephrology Dialysis Transplantation*, Volume 23, Issue 6, June 2008, Pages 2003–2009  
<https://doi.org/10.1093/ndt/gfm888>

<sup>5</sup>Björn Meijers, Christoph Metalidis, Thomas Vanhove, Ruben Poesen, Dirk Kuypers, Pieter Evenepoel, A noninferiority trial comparing a heparin-grafted membrane plus citrate-containing dialysate versus regional citrate anticoagulation: results of the CiTED study, *Nephrology Dialysis Transplantation*, Volume 32, Issue 4, April 2017, Pages 707–714  
<https://doi.org/10.1093/ndt/gfw461>